




INTUITIONLABS / RESEARCH & PRACTICAL GUIDANCE

CSV to CSA: Understanding FDA's New Validation Guidance

Build practical AI for pharma and biotech with IntuitionLabs. We help life-science teams turn complex information and workflows into useful software, governed knowledge systems and AI tools.

Published by IntuitionLabs · 2026-02-06 · computer software assurance · csv vs csa · fda guidance · risk-based validation · computer system validation · gxp compliance · software testing · quality systems · medical device software · pharma validation

Original article: <https://intuitionlabs.ai/articles/csv-to-csa-fda-validation-guidance>



In this article

1. [Executive Summary](#)
 2. [Introduction and Background](#)
 3. [Computer System Validation \(CSV\) – Established Principles and Burdensome Implementations](#)
 4. [Computer Software Assurance \(CSA\) – Risk-Based Assurance](#)
 5. [FDA's Current CSA Guidance \(Final\)](#)
 6. [CSA Implementation in Practice](#)
 7. [Evidence and Data on CSA Transition](#)
 8. [Illustrative Workflows](#)
 9. [Implications and Future Directions](#)
 10. [Conclusions](#)
-

Executive Summary

The U.S. Food and Drug Administration's (FDA) **Computer Software Assurance (CSA)** guidance clarifies risk-based assurance for software used in medical-device production and quality management systems. Risk-based validation predates CSA: FDA's 2002 guidance already bases validation coverage on software complexity and safety risk. The practical comparison is between burdensome, documentation-focused implementations of **Computer System Validation (CSV)** and CSA's more explicit recommendations for selecting assurance activities and evidence appropriate to intended use ([FDA CSA guidance](#); [FDA software validation guidance](#)).

This report traces the historical context, compares assurance practices, and discusses critical thinking, supplier evidence, testing methods, and lifecycle control. The scenarios are illustrative workflows rather than verified efficiency results. FDA's device **Quality Management System Regulation (QMSR)** has been effective since February 2, 2026 and incorporates ISO 13485:2016 ([FDA QMSR overview](#)). The discussion also considers accommodating AI within the guidance's medical-device scope. Pharmaceutical organizations should assess any application of these principles under applicable drug regulations; the February 2026 document is not a separate pharmaceutical CSA final guidance.

Introduction and Background

Today's life-sciences industries rely on a diverse, computerized ecosystem: from [manufacturing execution systems](#) and [laboratory information management systems](#), to cloud-based analytics and digital recordkeeping. Validation obligations depend on the system's intended use and the applicable regulatory framework. FDA's current CSA guidance addresses computers and automated data processing systems used in medical-device production or the quality management system ([FDA CSA guidance](#)). Historical context includes 21 CFR Part 11 for electronic records and FDA's 2002 "General Principles of Software Validation" (GPSV). That earlier guidance already recommended validation coverage based on software complexity and safety risk; CSV did not impose uniform exhaustive testing of every function ([FDA GPSV](#)).

Medical-device manufacturers have asked FDA for greater clarity about software validation expectations and for a more iterative, agile approach. FDA's current guidance describes those requests as part of its background ([FDA CSA guidance](#)). They provide context for clarifying assurance practices, rather than treating risk-based CSV as invalid.

Concurrently, digital manufacturing technologies have advanced. FDA describes automation, robotics, simulation, and other digital capabilities as opportunities to reduce error, optimize resources, and reduce patient risk ([FDA CSA guidance](#)). These technologies also make intended-use assessment, change control, and appropriate assurance evidence important when software is introduced or updated.

Against this backdrop, the FDA's "Case for Quality" initiative emerged (initiated ~2011, CDRH) to help modernize quality practices (www.fda.gov) (www.technologynetworks.com). Stakeholders identified CSV as a key *bottleneck* to innovation: companies noted duplication of effort (e.g. re-testing trusted vendor software), deterrence of automation, and focus on inspection evidence over product quality (www.technologynetworks.com). In response, FDA (particularly CDRH and CBER) and pharma leaders collaborated to develop a new paradigm. By 2015 a joint agency-industry team concluded that **Computer Software Assurance (CSA)** could deliver the same confidence with much lower burden (www.technologynetworks.com).

FDA released the CSA draft on September 13, 2022 ([FDA webinar transcript](#)). Its September 24, 2025 final device guidance was superseded on February 3, 2026 by "**Computer Software Assurance for Production and Quality Management System Software**" ([current FDA guidance](#)). The current document concerns medical-device production and quality management software. Pharmaceutical readers should evaluate risk-based assurance principles within their applicable drug regulatory framework, rather than rely on a purported pharmaceutical CSA final guidance. The rest of this report examines the practical changes within that scope.

Computer System Validation (CSV) – Established Principles and Burdensome Implementations

Validation Requirements and Risk: FDA's 2002 GPSV already recommended risk-based validation coverage. The current CSA guidance addresses medical-device software validation obligations within the QMSR framework and recommends assurance activities appropriate to intended use and risk ([FDA GPSV](#); [FDA CSA guidance](#)). Some implementations of CSV emphasized extensive protocols and traceability documentation, but that practice should not be mistaken for a historical FDA requirement to test all functions exhaustively.

Typical CSV Practice: In reality, CSV became synonymous with heavy documentation. Validation plans, requirement specs, design specs, traceability matrices, verification/validation protocols, change control, installation qualification (IQ), operational qualification (OQ), and performance qualification (PQ) — all were meticulously prepared and filed in massive binders. For example, one consultant noted that CSV often focused on creating paperwork *for auditors*, rather than truly assuring system behavior (www.technologynetworks.com) (ispe.org). Multiple industry sources lament that CSV's mantra was "*document it, qualify it, re-qualify it*", with limited attention to actual defect detection or critical thinking (www.technologynetworks.com) (www.bgosoftware.com).

Downsides of CSV: Over time the life-sciences industry and FDA recognized several systemic issues with the CSV approach:

- **Excessive Effort:** Organizations routinely spent months (or longer) on validation documentation. It was common to gather large validation teams and even external consultants for projects that might otherwise be routine software deployments. This effort often dwarfed the effort on actual testing or risk analysis (www.technologynetworks.com) (qmsdoc.com).
- **Focus on Paper Over Product:** Because auditors prioritized seeing documents, much of the validation work was “up-front” paperwork. In many cases, employees equated a bigger stack of paper with better quality, an inversion of the actual safety goal (www.technologynetworks.com) (ispe.org). FDA inspection findings revealed that even with hefty documentation, systems could still have undetected data-integrity, security, or configuration problems.
- **Duplication and Wasted Resources:** Whenever firms purchased a commercial off-the-shelf (COTS) software or cloud system, they often started validation from scratch. Many reported “no credit given” for reputable vendors’ own validation or certifications, forcing repeat testing. This duplication was seen as a significant waste (www.technologynetworks.com).
- **Inflexibility and Slow Updates:** The CSV model was largely sequential and stage-gated. Once a validation plan was set, changing course mid-project or embracing agile updates was difficult. Even low-risk changes often triggered full re-qualification cycles, delaying releases and frustrating IT staff.
- **Conservative Compliance Culture:** Life-sciences firms usually took an abundance-of-caution approach, aiming to eliminate all risk of inspection non-compliance. CSV became tightly bound to fear of 483 observations, which sometimes led to over-testing. As one industry writer notes, the pharmaceutical sector applied risk management sparingly, resulting in CSV practices that were “*inflexible, onerous, [and] a non-value-add*” (www.technologynetworks.com).

Across these years of feedback, both FDA and industry consensus formed that *validation practices needed modernization* if quality goals were to be met more effectively (www.fdl.org) (qmsdoc.com). The legacy CSV paradigm was addressing yesterday’s world; new technologies demanded new thinking.

Computer Software Assurance (CSA) – Risk-Based Assurance

Definition of CSA: FDA’s current medical-device guidance defines CSA as a risk-based approach for establishing and maintaining confidence that software is fit for its intended use. This builds on risk-based validation principles already present in GPSV ([FDA CSA guidance](http://FDA%20CSA%20guidance); [FDA GPSV](http://FDA%20GPSV)). A practical overview follows:

1. **Identify Intended Use:** Determine which features/functions of the software are actually used in production or quality processes (www.fdl.org) (www.jdsupra.com).
2. **Determine Risk:** Assess what could happen if each key function failed. This risk assessment is based on patient safety or product quality impact (www.fdl.org) (www.jdsupra.com).

3. **Define Assurance Activities:** Decide on verification, testing, and oversight activities proportional to each risk level. Higher-risk features require more rigorous evidence (possibly including vendor documentation), while lower-risk features can rely on lighter checks (www.fdli.org) (www.bgosoftware.com).
4. **Evidence and Records:** Document sufficient evidence to show confidence in each function “*but no more than necessary*.” Records should serve as a baseline for improvements, not excessive proof for its own sake (www.fdli.org) (www.fdli.org).

The assurance question is whether the software performs as intended and whether the evidence is appropriate to its identified risks. Maintaining a validated state includes lifecycle control ([FDA CSA guidance](#)).

Key Features of CSA: The CSA approach, as emphasized in FDA’s guidance and industry interpretations, includes these hallmarks:

- **Risk-Based Testing:** Select testing methods and rigor according to the intended use and analyzed risk. Scripted and unscripted methods are not restricted to particular risk categories; FDA notes that unscripted testing may be better suited even for high-process-risk functions ([FDA CSA guidance](#)).
- **Critical Thinking and Planning:** CSA stresses “*critical thinking*” early. Quality and compliance personnel engage in upfront analysis of risk and system purpose, rather than just executing pre-defined tests later (www.fdli.org) (ispe.org). This means identifying key failure modes, data integrity checks, and potential process effects as part of design reviews. It also encourages leveraging domain knowledge (e.g. real-world usage patterns) to target validation.
- **Vendor and Supplier Engagement:** Relevant supplier development, validation, and electronic evidence can form a starting point for assurance. Manufacturers determine what additional activities are needed for their intended use and risk. Supplier reputation or a GAMP category alone does not establish adequate assurance ([FDA CSA guidance](#)).
- **Proportionate Documentation:** Records should contain sufficient objective evidence that the assessed software performs as intended. Capture intended use, risk analysis, assurance activities, issues and their disposition, conclusions, who performed the assessment and when, and appropriate review and approval ([FDA CSA guidance, section V.A.6](#)).
- **Least-Burdensome Philosophy:** For the covered medical-device production and quality management software, FDA describes risk-based assurance as supporting efficient use of resources while promoting product quality ([FDA CSA guidance](#)).
- **Integrated Quality Oversight:** Unlike legacy CSV (where QA often entered a project late), CSA calls for QA-compliance roles to participate from the start. Quality/compliance teams collaborate on risk assessments, and remain involved throughout development and testing (www.fdli.org) (ispe.org). This shift helps catch gaps early (e.g., data integrity rules, audit trails), so they inform design rather than being “afterthoughts.” It also fosters a culture of quality across IT projects.

- **Lifecycle Focus (“Validated State”):** Manufacturers remain responsible for choosing assurance activities that maintain the software’s validated state. Assurance may include risk-based testing, monitoring, and relevant activities performed by suppliers, rather than a uniform full revalidation after every change ([FDA CSA guidance](#)).

CSV vs. CSA – Comparing Assurance Practices

Risk-based validation predates CSA. Table 1 contrasts burdensome implementations of CSV with the assurance practices clarified in FDA's current guidance for medical-device production and quality management software; it does not describe an FDA mandate for uniform exhaustive testing ([FDA GPSV](#); [FDA CSA guidance](#)).

Aspect	Burdensome CSV implementation	CSA recommendations for covered medical-device software
Approach	Uniform activities imposed by an organization's templates.	Assurance effort appropriate to intended use and identified device or process risk.
Testing Method	Scripts used by default, regardless of suitability.	Scripted, unscripted, or hybrid methods selected for suitability; unscripted testing can suit high-risk functions.
Documentation	Extensive paperwork and duplicated evidence.	Sufficient objective evidence of intended performance, with risk rationale, activities, issues, conclusions, and appropriate review.
Risk Assessment	Risk decisions may be obscured by fixed templates.	Analyze reasonably foreseeable failures and select commensurate assurance activities.
Vendor/Supplier	Relevant supplier work may be repeated unnecessarily.	Evaluate supplier evidence and determine additional activities needed for the manufacturer's intended use and risk.
Quality Involvement	Document approval may dominate the workflow.	Make assurance decisions and retain appropriate review and approval records.
Goal	Documentation volume becomes a proxy for assurance.	Establish confidence that software is fit for its intended use and maintain its validated state.
Outcome	Potential duplication and disproportionate effort.	A basis for efficient resource use; project-specific savings require measurement.

Table 1. Burdensome CSV implementations compared with FDA's clarified CSA practices. Both should be evaluated against established risk-based validation principles.

FDA’s Current CSA Guidance (Final)

FDA's “**Computer Software Assurance for Production and Quality Management System Software**”, issued February 3, 2026, supersedes the September 24, 2025 guidance. It provides recommendations for software used in **medical-device production or the quality management system** ([FDA guidance record](#); [FDA full guidance](#)). It does not provide design and development verification or validation recommendations for software functions that themselves meet the statutory definition of a medical device. It is not a separate pharmaceutical CSA final guidance.

Risk Framework and Assurance Activities

Determine intended use, identify reasonably foreseeable failures, and select assurance activities commensurate with device or process risk. For individual functions with different roles, document the corresponding risk rationale. Infrequent use or a peripheral label alone does not establish low risk; assess the consequences of failure and relevant controls ([FDA guidance, sections V.A.1–V.A.5](#)).

For higher risks, assurance generally entails greater rigor and more objective evidence. Scripted, unscripted, and hybrid testing are selected by suitability. FDA expressly notes that unscripted testing may be better suited even for high-process-risk functions ([FDA guidance, section V.A.4](#)).

Examples and Cloud or AI/ML Considerations

Appendix A includes examples involving nonconformance management, learning management, business intelligence, and a SaaS product lifecycle management system. The risk framework can apply to automation, data analytics, AI/ML tools, and cloud computing when used in the covered medical-device production or quality management context. These are assurance considerations, not technology credentials or blanket product approvals ([FDA guidance](#)).

Relationship to Prior Guidance and Regulations

The CSA guidance supplements GPSV and supersedes its Section 6 on automated process equipment and quality system software. It describes FDA's current thinking and is nonbinding; an alternative approach can be used if it satisfies applicable statutes and regulations ([FDA guidance](#)).

Device regulation has also changed: the **QMSR has been effective since February 2, 2026**, amending 21 CFR Part 820 and incorporating ISO 13485:2016 ([FDA QMSR overview](#)). Continuing legal obligations should therefore be distinguished from the amended regulatory text. CSA recommendations do not waive applicable legal requirements or establish blanket inspection acceptance.

CSA Implementation in Practice

Implementation starts with intended use and a documented function-specific risk rationale. The recommendations below concern the FDA guidance's medical-device production and quality management scope. Pharmaceutical organizations should evaluate any use of similar principles under their applicable drug regulations ([FDA guidance](#)).

- **Upfront Risk Assessment:** Identify reasonably foreseeable failures and their effects on product quality and safety. Consider configuration, security, data integrity, data transfer, and process controls. Avoid assigning a fixed risk level solely from a feature name or frequency of use ([FDA guidance, section V.A.2](#)).
- **Tailored Test Plans:** Select activities and evidence appropriate to intended use and analyzed risk. Higher risk generally calls for greater rigor, but does not mandate exclusively scripted testing. Scripted, unscripted, and hybrid approaches can be suitable across risk categories. Assess relevant supplier evidence and determine any additional assurance needed ([FDA guidance, sections V.A.4–V.A.5](#)).

- **Early QA Involvement:** Involve quality, IT, and process stakeholders in intended-use assessment and assurance planning. Retain appropriate review and approval of assurance records. These practices support clear responsibility throughout implementation without presuming that CSV historically excluded quality teams ([FDA guidance, section V.A.6](#)).
- **Use of Automation and Tools:** Digital retention, automated traceability, automated testing, and electronic capture of work can provide objective evidence and reduce manual documentation. This is not blanket FDA acceptance of a class of validation platforms as systems of record; assess the tool's intended use and applicable record requirements ([FDA guidance, section V.A.6](#)).
- **Supplier Evidence and Controls:** Use relevant supplier development and validation evidence as a starting point, then identify additional activities needed for intended use and risk. Assess supplier capabilities, configuration differences, data integrity and cybersecurity controls, and relevant quality agreements. Reputation, qualification, or a GAMP category alone is not a basis for minimal verification ([FDA guidance, section V.A.5](#)).
- **Documentation of Rationale:** Records should establish intended performance, rather than merely state that a function was assessed. Include intended use, risk analysis, assurance activities and results, issues and their resolution or risk justification, conclusions, the assessor and date, and appropriate review and approval ([FDA guidance, section V.A.6](#)).
- **Hybrid Testing Strategies:** Select scripted, unscripted, or hybrid testing according to suitability and analyzed risk. FDA's device guidance includes unscripted methods such as scenario testing, error guessing, and exploratory testing, and explains that unscripted testing can suit high-risk functions. These recommendations are not a regulator approval of pharmaceutical testing or peer-review methods ([FDA guidance, section V.A.4](#)).
- **Support for Lifecycle Changes:** Maintain a validated state across software changes. Determine assurance activities from the change's impact on intended use and risk, rather than assume that minor updates need only smoke testing or that every major update requires the same regression scope ([FDA guidance](#)).
- **Inspection Readiness:** Retain sufficient objective evidence of intended performance, including the risk rationale, assurance activities and results, issues and their disposition, conclusions, assessor and date, and appropriate review and approval. The guidance provides recommendations within its medical-device scope; it does not confer blanket inspection acceptance on a system or waive applicable legal documentation requirements ([FDA guidance, section V.A.6](#)).

Evidence and Data on CSA Transition

ISPE's workshop report describes more than 80 professionals discussing perceived barriers to CSA adoption in Orlando. Workshop participation is evidence of public discussion, rather than proof of a named company's deployment or an industry-wide adoption estimate ([ISPE workshop report](#)). The reported March and April 2024 webinar samples should not be treated as representative industry surveys.

FDA describes risk-based assurance as supporting efficient resource use and product quality ([FDA guidance](#)). This is a regulatory rationale, rather than a quantified result for the unnamed organizations previously described here. To evaluate benefits in a particular organization, measure validation effort, release timing, defects, and maintenance of the validated state before and after changes to assurance practices. Vendor reputation and public workshop engagement do not establish software performance or verified deployment outcomes.

Illustrative Workflows

The following scenarios are hypothetical. They do not report deployments, measured savings, product-quality outcomes, or inspection findings.

- **Medical Device Manufacturer:** *Acme Devices* is a fictional manufacturer evaluating an updated MES. Its team documents intended uses and reasonably foreseeable failures, considers relevant supplier evidence, and selects assurance activities suited to the resulting risks. Recipe control and a dashboard are assessed according to their actual effects; their names alone do not determine risk. Scripted, unscripted, or hybrid testing may be appropriate, with sufficient evidence of intended performance ([FDA guidance](#)).
- **Pharmaceutical Organization:** *HealthPharma Inc.* is a fictional organization evaluating a cloud LIMS. Its team assesses chain of custody, record integrity, exports, configuration, and relevant supplier evidence under applicable drug requirements. This scenario illustrates planning questions; it does not claim that the FDA device CSA guidance is a pharmaceutical final guidance or that checking only a few critical functions establishes compliance.

These workflows illustrate assurance decisions rather than demonstrate efficiency. The unnamed electronics and CRO accounts, named-company deployment inferences, savings percentages, and invented inspection outcomes have been removed because they were not supported by traceable primary case reports.

Implications and Future Directions

CSA recommendations sit within applicable quality-system obligations. Practical implications include:

- **Device Inspections:** FDA replaced QSIT on February 2, 2026 with the inspection process in its updated **Inspection of Medical Device Manufacturers Compliance Program: 7382.850** ([FDA QMSR overview](#)). Maintain evidence appropriate to intended use, risk, and applicable legal requirements. Do not infer blanket inspection acceptance across FDA, EMA, MHRA, or pharmaceutical systems from the device CSA guidance.
- **Quality Management System Framework:** The QMSR has been effective since February 2, 2026. It amends the device CGMP requirements of 21 CFR Part 820 and incorporates ISO 13485:2016 ([FDA QMSR overview](#)). FDA's current CSA guidance discusses software validation within this effective framework ([FDA CSA guidance](#)); the alignment is no longer forthcoming.

- **International Application:** Evaluate applicable requirements in each jurisdiction separately. The FDA CSA guidance describes recommendations for medical-device production and quality management software; it does not establish an approval by other regulators or displace pharmaceutical requirements ([FDA CSA guidance](#)).
- **Digital Transformation and Emerging Tech:** FDA's risk framework can be applied to automation, analytics, AI/ML tools, and cloud computing when used in the covered medical-device production or quality management context. Intended use and analyzed risk determine appropriate assurance activities; lower-risk functions still need adequate assurance. This does not establish a blanket approval of AI tools or a forecast of future pharmaceutical guidance ([FDA guidance](#)).
- **Vendor and Ecosystem Effects:** Software vendors and service providers will adapt to CSA. We anticipate more documentation and transparency from vendors about their own validation. Firms may pressure vendors to supply risk analyses and test results for their features (especially SaaS vendors). This could lead to “validation as a service” offerings, or standardized vendor evidence packets. Cloud platforms could incorporate CSA-compliant templates into their deployment tools, allowing plug-and-play risk alignment.
- **Culture and Training:** A lasting shift will be cultural. Quality personnel will need training in risk-based thinking and agile QA. CSA requires critical thinking skills: understanding process knowledge, evaluating failure modes, and justifying lean validation under audit. Companies might adopt cross-disciplinary teams of QA, engineering, and process owners. This could blur traditional QA/dev boundaries and encourage broader quality ownership across organizations.
- **Reduced Validation Costs and Time:** Although not yet systematically quantified, downstream effects are expected to include lower compliance costs. Resources saved on validation can be reallocated to continuous improvement, innovation, and improved quality controls. This aligns with the FDA's Case for Quality: ensure resource-intensive audits exist only where they most effectively protect patients ([qmsdoc.com](#)).
- **Future Guidance and Standards:** CSA guidance signals a **new era of validation**. The FDA and industry standards bodies are likely to issue further guidance elaborating modern computerized system practices. Just as CSA builds on GPSV and GAMP 5, we may see CSA-style updates in ICH guidelines (Q10, Q9) or new guidance on topics like Software as a Medical Device (SaMD) quality. The traction behind CSA suggests that other regulators (e.g. Health Canada, PMDA) may issue similar risk-based software validation guidance.

Conclusions

FDA's CSA guidance clarifies risk-based assurance for software used in medical-device production and quality management systems. Risk-based validation was already present in the 2002 GPSV; CSA should therefore be understood as a refinement of assurance practice, rather than a reversal of a historical FDA requirement for uniform exhaustive testing. Its practical focus is confidence in intended performance, with activities and evidence appropriate to identified risks ([FDA GPSV](#); [FDA CSA guidance](#)).

Implementation may require changes to training, tools, supplier oversight, and evidence practices. Teams should document intended use and risk, choose suitable assurance methods, and maintain adequate records. The illustrative workflows in this article do not establish quantified savings or inspection outcomes. Manufacturers remain responsible for maintaining a validated state, and the guidance is nonbinding: alternative approaches must satisfy applicable statutes and regulations ([FDA CSA guidance](#)). Pharmaceutical application requires consideration of the applicable drug framework; the February 2026 final guidance concerns medical devices.

Sources: The current regulatory analysis uses FDA's [February 2026 CSA guidance](#), its [guidance record](#), the [2002 GPSV](#), and FDA's [QMSR overview](#). Industry commentary retained elsewhere is attributed as commentary, rather than used to establish regulatory status or verified company outcomes.

Source links

Links cited in this article, in order of first appearance. Full addresses are provided for readers using extracted text.

1. <https://www.fda.gov/media/188844/download>
2. <https://www.fda.gov/media/73141/download>
3. <https://www.fda.gov/medical-devices/postmarket-requirements-devices/quality-management-system-regulation-qmsr>
4. <https://intuitionlabs.ai/articles/mes-ebr-pharma-compliance>
5. <https://intuitionlabs.ai/articles/lims-vs-elb-sdms-cds>
6. <https://www.fdi.org/2023/05/advancing-the-transition-to-computer-software-assurance/>
7. <https://www.technologynetworks.com/informatics/articles/computer-software-assurance-perfect-solution-or-confidence-trick-393177>
8. <https://www.fda.gov/media/163299/download>
9. <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/computer-software-assurance-production-and-quality-management-system-software>
10. <https://ispe.org/pharmaceutical-engineering/march-april-2024/computer-software-assurance-and-critical-thinking>
11. <https://www.bgosoftware.com/blog/csv-vs-csa-what-changes-what-doesnt-and-how-to-choose/>
12. <https://qmsdoc.com/2026/01/15/fda-issues-final-computer-software-assurance-guidance/>
13. <https://qmsdoc.com/2026/01/05/breaking-news-fda-issues-csa-guidance/>
14. <https://www.jdsupra.com/legalnews/fda-finalizes-computer-software-8878940/>
15. <https://ispe.org/pharmaceutical-engineering/september-october-2024/finding-assurance-computer-software-assurance>

THE PUBLISHER

About IntuitionLabs

IntuitionLabs is an AI consulting, custom software development and data engineering firm serving pharmaceutical, biotechnology, medical-device and other life-science organizations. We work with clinical, regulatory, medical-affairs, commercial, quality and IT teams to connect technology decisions with the work people need to accomplish.

AI consulting and adoption

Our AI enablement services cover readiness assessments, use-case selection, governance and policies, team workshops, adoption measurement and ongoing advisory support. We help organizations structure the information layer behind AI: source material, context, permissions and maintained knowledge that make generated answers useful and reviewable. Private LLM inference and hosted AI options support teams evaluating how to operate AI with appropriate control over their data and infrastructure.

Software, data and life-science workflows

IntuitionLabs develops custom software for pharma and biotech, integrates enterprise systems, and builds data engineering and business intelligence solutions. Areas of focus include AI agents, regulatory research, medical writing, medical affairs, CMC information, competitive intelligence and clinical-document workflows. Our eTMF intelligence work includes cross-system reconciliation and inspection-readiness support.

Enterprise platforms and regulated delivery

We provide Veeva services, application support, managed services, integrations and custom applications, alongside enterprise content work involving platforms such as Egnyte. For regulated workflows, our services include GxP enablement, computer-system validation and software development addressing 21 CFR Part 11 requirements. The applicable controls, validation responsibilities and acceptance criteria are defined for each engagement.

Work with IntuitionLabs

Explore [AI enablement](#), [pharma and biotech software development](#), [data engineering and BI](#), and [Veeva services](#). [Contact IntuitionLabs](#) to discuss your workflow, information sources and implementation needs.

IntuitionLabs publishes educational research to help life-science teams make informed technology decisions. Coverage of a product or organization does not imply a client relationship, endorsement or partnership.

Website: <https://intuitionlabs.ai>

This article is educational. Verify consequential medical, legal, financial, regulatory and technical decisions with appropriate professionals and the cited primary sources. Company services are described separately from the article's research findings.